



Empore™

Radium RAD Disks

Method Summary: Test Method RA-195

Rapid Determination of Radium-228 in Water by Elution of Ingrown Actinium-228 from Empore™ Radium RAD Disks

Product Description

Empore™ Radium RAD Disks provide an efficient alternative to conventional radiochemical sample preparation methods that use wet chemistry or packed columns. A proprietary process is used to entrap radium selective adsorbent particles into a matrix of inert PTFE to create a mechanically stable sorbent disk. The disks can be used for purification and concentration of radium from aqueous samples.

Empore radium RAD disks provide a sample prep solution for large volume aqueous samples and the disk format provides a large surface area for sorbent/sample contact.

Scope and Application

This method uses Empore radium RAD disks to concentrate and purify radium from water samples, which results in significant reductions in sample preparation time over EPA Method 904.0. The method assumes quantitative extraction of soluble radium from the water sample, and at present, no correction for chemical yield has been included.

Method Overview

An acidified sample is passed through an Empore radium RAD disk to quantitatively extract soluble radium in the sample. After completion of a suitable ingrowth period, actinium-228 is eluted with 0.5 M HNO₃ and analyzed using a low background gas flow proportional counter.

Interferences

Sr²⁺ is not retained by the disk. Quantitative radium recovery (>95%) has been shown in one liter samples with the following individual concentrations of potential interfering ions:

Mg ²⁺	10,000 mg/L	Sr ²⁺	2.0 mg/L
Ca ²⁺	10,000 mg/L	Na ⁺	1,000 mg/L
Pb ²⁺	1.0 mg/L	K ⁺	10 mg/L
Ba ²⁺	1.0 mg/L		

'True' interference concentrations may be even higher than those tested.

Safety

When using solvents or other chemicals, be sure to read and follow the manufacturer's precautions, directions for use and disposal procedures.

Apparatus and Materials

- 47 mm Empore radium RAD disks
- 47 mm single or multiple station vacuum manifold
- Beta detector (low background gas flow proportional counter)
- 50 mm planchets

Reagents

Nitric Acid (HNO₃): 16 M, 2 M, and 0.5 M

Sample Collection, Preservation, and Handling

Samples may be acidified to pH < 2 with HNO₃ prior to sub-sampling.

Procedure

1. Make the sample 2 M in HNO₃ by adding 126 mL concentrated HNO₃ per liter of sample being processed. If membrane clogging is anticipated, the sample may be prefiltered through a 0.45 μ filter.
2. Mount an Empore radium RAD disk in the filter support.
3. Wash the disk by drawing 20 mL of 2 M HNO₃ through the disk with a gentle vacuum. Do not allow the disk to go dry.
4. Extract the sample by pulling it through the disk at a nominal flow rate of 50 mL/min (about 20 minutes per liter).
5. Rinse the disk with 20 mL of 2 M HNO₃. The end of this rinse is recorded as the start of the actinium-228 ingrowth.
6. Remove the disk from the manifold and store it in a 50 mm Petri dish for 14-28 days to allow for ingrowth of actinium-228 and decay of radium-224.
7. Following storage, remount the disk in the filter support as described in step 2.
8. Elute actinium with 15 mL of 0.5 M HNO₃.
9. Mount the solution for gas flow proportional counting by evaporation on a planchet, by coprecipitation on yttrium oxalate (as described in EPA method 904.0), or by fluoride microprecipitation with neodymium or other suitable lanthanide.
10. Count the source as quickly as possible due to the very short half-life of actinium-228 (T_{1/2}=6.15 hours). The method is calibrated against standard actinium-228 mounted in a geometry equivalent to that of the sample.

Reference

Environmental Protection Agency, "Radium-228 in Drinking Water – Method 904.0", Section 8, *Prescribed Procedures for Measurement of Radioactivity in Drinking Water*, EPA-600/4-80-032, August, 1980.

Note: Empore Solid Phase Extraction Products are intended for solid phase extraction during scientific research only. These products are not intended for use in medical devices or in assessment and treatment of clinical patients.



465 Limestone Rd, Oxford, PA 19363
Toll Free: 1-800-541-6593

CDS Analytical and Empore are trademarks of CDS Analytical Company.
© CDS Analytical LLC 2018.
All Rights Reserved.
Printed in the USA.
GN-18 10/23 5K

Important Notice

CDS Analytical MAKES NO WARRANTIES, EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. Since a variety of factors can affect the use and performance of a CDS Analytical product in a particular application, some of which are uniquely within the user's knowledge and control, user is responsible for determining whether or not the CDS Analytical product is fit for a particular purpose and suitable for user's method of application.

Warranty

Seller warrants its equipment against defects in workmanship and material for a period of 12 months from date of shipment from the factory under normal use and service and otherwise when such equipment is used in accordance with instructions furnished by Seller and for purposes disclosed in writing at the time of purchase, if any. Any unauthorized alteration or modification of the equipment by Buyer will void this warranty. Seller's liability under this warranty shall be limited to the replacement or repair, F.O.B., point of manufacture, of any defective equipment or part which, having been returned to the factory, transportation charges prepaid, has been inspected and determined by Seller to be defective. THIS WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EITHER EXPRESSED OR IMPLIED, AS TO DESCRIPTION, QUALITY, MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE OR USE, OR ANY OTHER MATTER. Under no circumstances shall Seller be liable to Buyer or any third party for any loss of profits or other direct or indirect costs, expenses, losses or consequential damages arising out of or as a result of any defects in or failure of its products or any part or parts thereof or arising out of or as a result of parts or components incorporated in Seller's equipment but not supplied by the Seller.

Technical Information and Data

Technical information and data, recommendations, and other statements provided by CDS Analytical are based on information, tests, or experience which CDS Analytical believes to be reliable, but the accuracy or completeness of such information is not guaranteed. Such technical information and data are intended for persons with knowledge and technical skills sufficient to assess and apply their own informed judgment to the information. No license under any CDS Analytical or third party intellectual property rights is granted or implied with this information.

Safety and Disposal

Always follow safe handling practices while processing Empore products as applicable to your use. Always wear appropriate personal protective equipment while processing Empore products as applicable to your use. Dispose of used Empore products per requirements of your process. Alteration or addition of other materials to Empore products may require unique disposal protocols. Always dispose of used product in accordance with federal, state and local requirements.

Safety Information

WARNING:	Indicates a hazardous situation which, if not avoided, could result in death or serious injury.
WARNING:	
To reduce the risks associated with environmental contamination, handling radiochemical materials or false negative results:	
<ul style="list-style-type: none"> • Read, understand and follow all recommendations in the product MSDS and Instructions for Use. 	
To reduce the risks associated with radiochemical materials:	
<ul style="list-style-type: none"> • Always wear the proper PPE (Personal Protection Equipment) when handling used product. 	
To reduce the risks associated with environmental contamination:	
<ul style="list-style-type: none"> • Dispose of used disks according to standard radiochemical procedures and regulations. Unused disks may be disposed of in accordance with standard procedure for laboratory waste. 	